

Certificate



Certificate no.

CU 72110453 01

License Holder:

Diagast SAS
Parc Eurasanté
251 avenue Eugène Avinée - B.P.9

59374 Loos Cedex
France

Manufacturing Plant:

Diagast SAS
Parc Eurasanté
251 avenue Eugène Avinée - B.P.9

59374 Loos Cedex
France

Test report no.: USA-DS 31180362 001

Client Reference: Olivier Fleurie

Tested to:

UL 61010-1:2004 R10.08
IEC 61010-1:2001
IEC 61010-2-081:2001+A1
IEC 61010-2-101:2002
CAN/CSA-C22.2 NO. 61010-1-04+GI1 (R2009)
CAN/CSA-C22.2 No. 61010.2.081-04 (R2009)
CAN/CSA-C22.2 No. 61010.2.101-04 (R2009)

Certified Product: Fully Automated Blood Grouping System

License Fee - Units

Model Designation: QWALYS 3 (Type Reference 90804, 90854)

7

Rated Voltage: AC 100-120/200-240V, 50/60Hz
Rated Power: 400VA
Protection Class: I

Appendix: 1, 1-4

7

Licensed Test mark:



Signature

Dipl.-Ing. M. Glagla
QA Certification Officer

Date of Issue

(day/mo/yr)

01/06/2011