



Ref. Certif. No.

DE 2-025907

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT (IECEE) CB SCHEME

CB TEST CERTIFICATE

Product

Fully automated blood grouping system

Name and address of the applicant

DIAGAST SAS
251 AVENUE EUGENE AVINEE
59120 LOOS, France

Name and address of the manufacturer

DIAGAST SAS
251 AVENUE EUGENE AVINEE
59120 LOOS, France

Name and address of the factory

DIAGAST SAS
251 AVENUE EUGENE AVINEE
59120 LOOS, France

Ratings and principal characteristics

AC 100-120V / 220-240V; 50/60Hz; 1200VA; class I

Trademark (if any)

DIAGAST

Customer's Testing Facility (CTF) Stage used

CTF stage 1

Model / Type Ref.

QWALYS 3 (Type reference 90854)

Additional information (if necessary may also be reported on page 2)

A sample of the product was tested and found to be in conformity with

IEC 61010-1:2010
IEC 61010-2-081:2015
IEC 61010-2-101:2015
for national deviations see test report

As shown in the Test Report Ref. No. which forms part of this Certificate

60245742 001

This CB Test Certificate is issued by the National Certification Body



TÜV Rheinland LGA Products GmbH
Tillystraße 2 · 90431 Nürnberg, Germany
Phone + 49 221 806-1371
Fax + 49 221 806-3935
Mail: cert-validity@de.tuv.com
Web: www.tuv.com



Date: 18.09.2019

Signature: Dipl.-Ing. A. ...